Criterion	**Minimal Requirement**
Spatial Selectivity: Radial sensitivity distribution (farfield) Spatial Resolution	**FWHM ≤ 40°** High demands at lymph nodes in head-, neck-, supraclavicular region**:** **FWHM ≤ 15mm** Normal requirements at extremeties, axilla, groin: **FWHM ≤ 25mm**
Sensitivity	**≥ 5000 cps/MBq**
Shielding	**≤ 0,1 % of maximum system sensitivity**
Energy selection	Compton/photopeak discrimination; check of energy selection possible
Display Acoustic digital **or** analogue	 Good correlation between measurement signal and tone digital: continuous display with adjustable measurement time interval analogue: suitable measurement interval with adjustable time constant